HCL001\_HCL002\_HCL003\_PI\_V008



# **HPV/p16** Analyte Control<sup>DR</sup>

Cell line controls for immunohistochemistry and in situ hybridization.

## Research Use Only

## PRODUCT AVAILABILITY

| Product Code | Product Description      |
|--------------|--------------------------|
| HCL001       | X2 Cut slides            |
| HCL002       | X5 Cut slides            |
| HCL003       | X1 Cell microarray block |

#### **APPLICATION**

This product is suitable for use in immunohistochemistry and in situ hybridization.


#### **MATERIALS**

Four formalin fixed paraffin embedded cell lines. Three of which contain incorporated HPV 16 and 18 viral genes.


Cell line A: Human breast adenocarcinoma
Cell line B: Human cervical squamous cell carcinoma
Cell line C: Human cervical adenocarcinoma
Cell line D: Human epidermoid carcinoma

Cells are fixed in 10% neutral buffered formalin and paraffin wax

Sections are cut at  $4\mu m$ , mounted on positively charged slides and baked overnight at  $37^{\circ}C$ .



Cell microarrays (CMA) contain cores that are 1.5-2mm in diameter and 3-3.5mm in length. It is possible to obtain over 300 sections depending on thickness.



## **Expression Profile**

| Cell<br>Line | HPV Gene<br>Copy         | mRNA E6/E7<br>Copy | p16 Expression |
|--------------|--------------------------|--------------------|----------------|
| A            | Negative                 | Negative           | Negative       |
| В            | Low (1-2<br>copies/cell) | Low                | High           |
| С            | Medium                   | High               | High           |
| D            | High                     | High               | High           |

#### Storage and Handling

Store at 2-8°C. Do not freeze (for expiration date please see the product label)

# **WARNINGS AND PRECAUTIONS**

- 1. The product is intended for research use only.
- It is the responsibility of the end user to determine suitability with their reagents and procedures within their laboratory.
- Do not use after expiration date printed on product labels.
   The user must validate any storage conditions other than those specified in the package insert.

## **TROUBLE SHOOTING**

For further help please feel free to contact HistoCyte Laboratories Ltd at info@histocyte.com or call +44 (0)191 603 1007.



For updates and additional product information please visit: www.HistoCyte.com